


Patient Information

Patient ID : 12345678
Name : Lastname, Firstname
 Birth Date : 09-Apr-1976
 Professional : Yourname
 Today's Date : 09-Aug-2020, 16:24:13

Street : Street
 City : City
 State/Prov. : Province A1B 2C3
 Country : Canada
 Phone : (519) 123-4567

LEFT EAR

Speech SPLogram



HEARING AID RECOMMENDATION

LEFT EAR

Selection Method : DSL [i/o]

HEARING AID

Style : BTE
 Make :
 Model :
 Serial # :

EARMOLD

Type : SHELL
 Tube : NORMAL #13
 Bore : STANDARD #13
 Vent : PRESSURE 1MM
 Material : NEWSIL

OTHER

Transducer : ER3
 HL to SPL : RECD
 HA Style : BTE
 Circuit : WDRC (fixed CR)
 RE to 2cc : Measured

OTHER

Speech : Cox/Moore
 Compr. Thresh : 50
 Loudness : Predicted
 Max. Out : Measured

	<u>.25</u>	<u>.50</u>	<u>.75</u>	<u>1.0</u>	<u>1.5</u>	<u>2.0</u>	<u>3.0</u>	<u>4.0</u>	<u>6.0</u>
SSPL-90	105	109	116	115	121	126	118	115	97
Full-On Gain (Reserve 10 dB)	29	28	32	33	36	43	42	42	35
User Gain (Input 65 dB)	19	18	22	23	26	33	32	32	25
Comp. Ratio	2.0	1.8	1.8	2.0	2.0	2.2	2.5	2.8	6.3